

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO. FILING DATE FIRST NAMED INVENTOR ATTORNEY DOCKET NO. CONFIRMATION NO. 10/712,374 11/14/2003 Ghaleb A. Sater P1384 9586 28390 7590 09/07/2005 **EXAMINER** MEDTRONIC VASCULAR, INC. STIGELL, THEODORE J IP LEGAL DEPARTMENT ART UNIT PAPER NUMBER 3576 UNOCAL PLACE SANTA ROSA, CA 95403 3763

DATE MAILED: 09/07/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

		<u> </u>
Office Action Summary	Application No.	Applicant(s)
	10/712,374	SATER, GHALEB A.
	Examiner	Art Unit
	Theodore J. Stigell	3763
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply		
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).		
Status		
1)⊠ Responsive to communication(s) filed on 14 N	lovember 2003	
,	s action is non-final.	
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.		
Disposition of Claims		
4) Claim(s) 1-16 is/are pending in the application 4a) Of the above claim(s) is/are withdra 5) Claim(s) is/are allowed. 6) Claim(s) 1-16 is/are rejected. 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/o Application Papers 9) The specification is objected to by the Examine 10) The drawing(s) filed on 14 November 2003 is/a Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the E	er. are: a) accepted or b) objected drawing(s) be held in abeyance. Section is required if the drawing(s) is objected or by objection is required if the drawing(s) is objection is required if the drawing(s) is objection is required.	e 37 CFR 1.85(a). jected to. See 37 CFR 1.121(d).
Priority under 35 U.S.C. § 119		
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 		
Attachment(s) 1) ☑ Notice of References Cited (PTO-892) 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) ☑ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08 Paper No(s)/Mail Date 121 2004	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:	

Application/Control Number: 10/712,374

Art Unit: 3763

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-2,4-8,10-11, and 13-15 are rejected under 35 U.S.C. 102(b) as being anticipated by Jayaraman (6,312,406). Jayaraman clearly shows a catheter with a first shaft (102), a second catheter shaft (106) having a lumen for a guidewire, the second shaft including a plurality of guidewire clips (108) for holding the guidewire along the first shaft. The plurality of clips includes proximal and distal clips and at least two intermediate clips disposed between the proximal and distal clips. See Figure 3A. The clips can be arranged along the proximal portion of the first shaft (Figure 3D) or the distal portion of the first shaft (3C). The catheter can be in the form of a balloon catheter with an inflation lumen (102) and a balloon (104) at the distal end in fluid communication with the lumen. The catheter can also be used in a stent system with a balloon and a stent included. See column 4, lines 8-10. The catheter can also be used an aspiration catheter as seen in Figure 9 and 10 with holes (204,208) in fluid communication with the aspiration lumen.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

Art Unit: 3763

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

- 1. Determining the scope and contents of the prior art.
- 2. Ascertaining the differences between the prior art and the claims at issue.
- 3. Resolving the level of ordinary skill in the pertinent art.
- 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 3,9,12, and 16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Jayaraman (6,312,406) in view of Gharibadeh et al. (5,458,613). Jayaraman discloses a catheter that includes all of the limitations as recited in claims 1, 7, 10, and 14. Jayaraman does not teach to add slits to the outer surface of the guidewire clips to allow for transverse loading and unloading. Gharibadeh et al. disclose a rapid exchange catheter that allows for exchange of a catheter during an intraluminal procedure. Gharibadeh et al. disclose a catheter with a slit (26) on the outside wall of the guidewire lumen to allow for loading and unloading of the guidewire along any portion of the catheter. Therefore, it would be obvious to one of ordinary skill in the art at the time of the invention to modify the catheter of Jayaraman with the guidewire slits of Gharibadeh et al. to make a catheter that can be exchanged in a more rapid fashion.

Conclusion

Application/Control Number: 10/712,374 Page 4

Art Unit: 3763

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Theodore J. Stigell whose telephone number is 571-272-8759. The examiner can normally be reached on M-F 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Theodore Stigell

Header Hijell

NICHOLAS D. LUCCHESI SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700